

## ordering information

### Micro

MicroGalaxy CO<sub>2</sub> Incubator for 0.2-20% CO<sub>2</sub>

order code

Voltage 220/240V 50/60Hz  
100/120V 50/60Hz

015-700

015-701

**No charge option (factory fitted)**

Left Opening Door (hinged at right)

015-727

**Model options (factory fitted)**

O<sub>2</sub> Control 1 – 19%

015-711

Access port, 25mm diameter

015-719

RS 232 port \*

015-720R

Inner Glass Door \*

015-722

BMS Relay Contact Alarm \*

015-723

Key Lock on Door

015-733

**\* Can be retro-fitted by Service Personnel**



## accessories

order code

Galaxy IR CO<sub>2</sub> Analyser

010-100

Galaxy Combined CO<sub>2</sub> & O<sub>2</sub> analyser

010-1000

Perforated Shelves

015-104

Non-perforated Shelves

015-105

Mounting Bracket – Single Unit

015-116

GalaxyLog V1.0 PC based data collection software and RS232 port

015-117

Mounting Bracket – Two Units Stacked

015-118

Multi-position Shelf Racks (4 positions)

015-121

Automatic CO<sub>2</sub> Change-over Unit

170-101

CO<sub>2</sub> Two Stage Regulator

170-102

CO<sub>2</sub> Filter (supplied in pairs)

170-103

In-line Pressure Regulator

170-110

CO<sub>2</sub> Gas Analyser complete with Testing Tubes

170-113

10 Spare Gas Analyser Testing Tubes

170-114

12V DC to 230V AC Inverter (UK Only)

015-122

RO Statoh recommends that an in-line pressure regulator (part no.170-110) is fitted to all gas supply lines whenever O<sub>2</sub> control is employed.

Warnings with options

Chamber conditions

Multi-position shelf rack

Smooth Interior  
Radii 14 mm  
(per unit) or  
(total)

Could be used



MicroGalaxy CO<sub>2</sub> Incubators.  
Small • Safe • Secure

14 Litres of  
Personal Space

C&M SCIENTIFIC  
Unit 4, Lindsay Square, Deans Industrial Estate, Livingston, West  
Lothian, EH54 8RL

